



SAMALESWARI REGIONAL CO-OPERATIVE MILK PRODUCERS' UNION LTD.

Regd. No.6/SM of 6th March 76

SAKHIPARA, SARLA CANAL ROAD, DHANKAUDA, DIST-SAMBALPUR (ODISHA) 768 006

Phone : (0663) 2548590, Fax : (0663) 2548520

Website - www.samul.in

E-mail : omfedsrcmpul@gmail.com / omfedsrcmpul@yahoo.com

Ref. No.: SRMU/ 270

Dated:- 22.03.2018

QUOTATION CALL NOTICE


Sub : INVITATION FOR QUOTATIONS FOR SUPPLY OF **ELECTRONICS MILKO TESTER.**

1. You are invited to submit your most competitive quotation for the following goods:-

SI No	Brief Description of the Goods	Specifications & Make & Model	Quantity	Delivery Period	Place of Delivery	Installation Requirement if any
1	Electronics Milko Tester without Handle	Any reputed make <i>Detail Technical Specification enclosed.</i>	40 Nos.	Within 30 days from the date of issue of Supply Order	Milk Chilling Plant, Boxma	Yes
2	Electronics Milko Tester with Handle	Any reputed make <i>Detail Technical Specification enclosed.</i>	40 Nos.	Within 30 days from the date of issue of Supply Order	Milk Chilling Plant, Boxma	Yes

* Where ISI certification marked goods are available in market, purchase should generally be limited to goods with those or equivalent marking only.

1. Interested Manufacturer/Supplier/Distributor may quote their rate indicating price, Tax,, Terms of Rate, Guarantee etc to the above address. **The detail technical specification, features of the Goods should be clearly mentioned in the quotation.** Boucher/Leaflet of the goods be attached with the quotation.
2. GST Extra.
3. Quotation should remain valid for a period not less than 60 days after the deadline date specified for submission.
4. Quotation should conform to the terms and conditions, and specifications.
5. Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject any or all quotations at any time prior to the award of contract.
6. Payment shall be made within 30 days of delivery of the goods.
7. Normal commercial warranty/ guarantee shall be applicable to the supplied goods.
8. Term of supply i.e regarding Packing & forwarding, Insurance, Delivery period and term of delivery should be clearly mentioned.
9. Quotation should reach latest by 15:00 hrs. on **dt.06.04.2018** for our consideration.
10. All disputes are subject to Sambalpur jurisdiction.


22/3/18
General Manager

To :



SAMALESWARI REGIONAL CO-OPERATIVE MILK PRODUCERS' UNION LTD.

Regd. No.6/SM of 6th March 76

SAKHIPARA, SARLA CANAL ROAD, DHANKAUDA, DIST-SAMBALPUR (ODISHA) 768 006

Phone : (0663) 2548590, Fax : (0663) 2548520

Website – www.samul.in

E-mail : omfedsrcompul@gmail.com/ omfedsrcompul@yahoo.com

TECHNICAL SPECIFICATION OF ELECTRONICS MILKO TESTER

Measuring Range	:	Up to 13%
Capacity	:	120 – 150 samples per hours
Accuracy (sd)	:	0 – 5% fat : 0.06% 5 – 8% fat : 0.10% 8 – 13% fat : 0.20%
Repeatability (sd)	:	0 – 5% fat : 03% 5 – 8% fat : 04% 8 – 13% fat : 0.08%
Sample Volume	:	0.5 ml/test
Diluent Volume	:	6.5 ml./test
Power supply	:	AC : 220 – 240 V Maximum + 10% Minimum 15% DC : 12 V, 6A, Max, 6V, Min 10.5 A fully charged 120 amp/hr. battery Will last for at least 10 hours operation.
Ambient temperature	:	5 – 45° C

SPECIFICATION OF VOLTAGE STABILISER

Automatic Voltage Stabiliser for Milko Testers

Capacity	:	500 VA
Input	:	60 V – 260 V
Output	:	200 V – 240 V AC with high/low Cut Off & Delay on.

The provision of embossed must be in Stabilizer.